

Hand Sanitiser / Hand Rub

We are seeking hand sanitiser produced in Australia including Therapeutic Goods Administration (TGA) approved hand sanitiser.

To enable additional supply of hand sanitiser, the TGA is also undertaking ongoing reviews regarding product regulations for hand sanitiser.

SCOPE

We are looking for the following types of hand sanitisers:

Non-alcohol (for use outside hospitals)

- With chlorhexidine
- Without chlorhexidine

Alcohol-based

- With chlorhexidine
- Without chlorhexidine



SPECIFICATIONS

At a minimum, all products must meet the following standards and testing requirements:

- AS/NZS ISO 9001:2008
- EN1276, prEN12054 and EN1500
- 21CFR 333.470
- ASTM E-1174

In addition to the minimum standards listed above, alcohol-based hand rubs must have an alcohol concentration between 60% to 90%.

Alcohol based hand sanitiser for use within hospitals must be TGA approved