

Gloves

Australia requires manufacturers of examination and surgical gloves. We are seeking Australian based manufacturers to set up production to support NSW.



We are looking for the following types of gloves:

Non-sterile

SCOPE

Sterile

- Latex
- Surgical, non-latex, accelerator-free

SPECIFICATIONS

At a minimum, the products must meet the following standards:

- ACORN Standards 2016
- ISO 2859, ISO 13485
- EN 455: Parts 1 to 4, EN 374: Parts 1 to 3, EN 420, EN 16523-1
- ISO 11193-1:2008: Parts 1 & 2
- ASTM D3578, D6319

- ACORN Standards 2016
- ISO 2859
- ISO 10282:2014 MOD, ISO 11137-1, ISO 13485, ISO 14001, ISO 9001
- ASTM D3577
- EN 455: Parts 1 to 4, EN 374: Parts 1 to 3, EN 420