

# Masks

---

We are seeking Australian based manufacturers with the ability to set up production of a range of surgical masks to support NSW.



## SCOPE

We are looking for the following types of masks:

### P2

## SPECIFICATIONS

At a minimum, the products must meet the following standards:

- AS/NZS 1716:2012, AS/NZS 1715:2009
- Fluid resistance: Level 1 - 80mm Hg , Level 2 - 120mm Hg , Level 3 - 160mm Hg

### N95

- AS/NZS 1716:2012, AS/NZS 1715:2009
- National Personal Protective Technology Laboratory (NPPTL) tested
- National Institute for Occupational Safety and Health (NIOSH) certified
- Fluid resistance: Level 1 - 80mm Hg , Level 2 - 120mm Hg , Level 3 - 160mm Hg

### Surgical masks

- AS /NZ4381:2015,
- Certified Level 1 - 3 barriers medical face masks are evaluated for resistance to penetration by synthetic blood at minimum velocity specified for bacterial filtration efficiency and differential pressure
- ASTM F2101-14 or EN 14683:2014 ASTM F1862 / F1862M-13 or ISO 22609 certified